

Certified Calibration Fluid  
 Batch No. 0008714433  
 Part No. CALI0000005

Grace Instrument Company  
 Houston, TX 77043  
 (713) 783-1560

20 cP

°C	Centipoise	°C	Centipoise	°C	Centipoise	°C	Centipoise
20.0	22.60	25.0	20.55	30.0	18.74	35.0	17.16
20.1	22.56	25.1	20.51	30.1	18.71	35.1	17.13
20.2	22.52	25.2	20.48	30.2	18.68	35.2	17.10
20.3	22.48	25.3	20.44	30.3	18.65	35.3	17.08
20.4	22.44	25.4	20.41	30.4	18.61	35.4	17.05
20.5	22.40	25.5	20.37	30.5	18.58	35.5	17.02
20.6	22.35	25.6	20.33	30.6	18.55	35.6	16.99
20.7	22.31	25.7	20.30	30.7	18.52	35.7	16.96
20.8	22.27	25.8	20.26	30.8	18.49	35.8	16.94
20.9	22.23	25.9	20.22	30.9	18.46	35.9	16.91
21.0	22.19	26.0	20.19	31.0	18.42	36.0	16.88
21.1	22.15	26.1	20.15	31.1	18.39	36.1	16.85
21.2	22.11	26.2	20.12	31.2	18.36	36.2	16.82
21.3	22.07	26.3	20.08	31.3	18.33	36.3	16.80
21.4	22.03	26.4	20.04	31.4	18.30	36.4	16.77
21.5	21.99	26.5	20.01	31.5	18.27	36.5	16.74
21.6	21.94	26.6	19.97	31.6	18.23	36.6	16.71
21.7	21.90	26.7	19.93	31.7	18.20	36.7	16.68
21.8	21.86	26.8	19.90	31.8	18.17	36.8	16.66
21.9	21.82	26.9	19.86	31.9	18.14	36.9	16.63
22.0	21.78	27.0	19.83	32.0	18.11	37.0	16.60
22.1	21.74	27.1	19.79	32.1	18.08	37.1	16.57
22.2	21.70	27.2	19.75	32.2	18.04	37.2	16.54
22.3	21.66	27.3	19.72	32.3	18.01	37.3	16.52
22.4	21.62	27.4	19.68	32.4	17.98	37.4	16.49
22.5	21.58	27.5	19.65	32.5	17.95	37.5	16.46
22.6	21.53	27.6	19.61	32.6	17.92	37.6	16.43
22.7	21.49	27.7	19.57	32.7	17.89	37.7	16.40
22.8	21.45	27.8	19.54	32.8	17.86	37.8	16.38
22.9	21.41	27.9	19.50	32.9	17.82	37.9	16.35
23.0	21.37	28.0	19.46	33.0	17.79	38.0	16.32
23.1	21.33	28.1	19.43	33.1	17.76	38.1	16.29
23.2	21.29	28.2	19.39	33.2	17.73	38.2	16.26
23.3	21.25	28.3	19.36	33.3	17.70	38.3	16.24
23.4	21.21	28.4	19.32	33.4	17.67	38.4	16.21
23.5	21.17	28.5	19.28	33.5	17.63	38.5	16.18
23.6	21.12	28.6	19.25	33.6	17.60	38.6	16.15
23.7	21.08	28.7	19.21	33.7	17.57	38.7	16.12
23.8	21.04	28.8	19.17	33.8	17.54	38.8	16.10
23.9	21.00	28.9	19.14	33.9	17.51	38.9	16.07
24.0	20.96	29.0	19.10	34.0	17.48	39.0	16.04
24.1	20.92	29.1	19.07	34.1	17.44	39.1	16.01
24.2	20.88	29.2	19.03	34.2	17.41	39.2	15.98
24.3	20.84	29.3	18.99	34.3	17.38	39.3	15.96
24.4	20.80	29.4	18.96	34.4	17.35	39.4	15.93
24.5	20.76	29.5	18.92	34.5	17.32	39.5	15.90
24.6	20.71	29.6	18.88	34.6	17.29	39.6	15.87
24.7	20.67	29.7	18.85	34.7	17.25	39.7	15.84
24.8	20.63	29.8	18.81	34.8	17.22	39.8	15.82
24.9	20.59	29.9	18.78	34.9	17.19	39.9	15.79
						40.0	15.76